



नेपाल सरकार

स्वास्थ्य तथा जनसंख्या मन्त्रालय
कोशी अस्पताल
बिराटनगर, मोरङ

मिति-२०८२। १०। २३ गते

बिषय: दररेट उपलब्ध गराइदिने सम्बन्धमा।

श्री

प्रस्तुत बिषयमा चालु आ.व.०८२।०८३ को बजेट तथा कार्यक्रम अनुसार यस अस्पतालको लागि तपशीलको सलग्न बमोजिमको Infant Warmer with sensor खरिद गर्नुपर्ने भएकोले लागत अनुमान तयार पार्ने उदेश्यका लागि सम्बन्धीत फर्मले उपलब्ध गराउन सक्ने दररेट तथा फर्मका आवश्यक कागजात सहित ७(सात) दिन भित्र email-info@koshihospital.gov.np मा उपलब्ध गराई सहयोग गरिदिनु हुन अनुरोध छ ।

दररेट पेश गर्नुपर्ने सामानको विवरण-

SN	Item Name and Description	Specification	Name of Company/Country Brand/Model	Quantity	Unit	Per unit Rate	Vat	Total
1	Infant Warmer with sensor	सलग्न छ		1	set			

Summarized Technical Specification of Infant Warmer

1. It shall be microprocessor controlled Infant warmer with manual and servo options.
2. Height of the bassinet must be adjustable as per the user's convenience and can be operated from both sides.
3. The heating element must be ceramic or quartz infrared or similar technology with uniform heat radiation or better. Whole heater unit should swivels away from bed to make room for an x-ray unit.

4. Bed tilting: ± 12 degree (approx) and Bed rotation: ± 60 degree (approx)
5. The unit should be supplied with superior quality Mattress to provide maximum patient safety and comfort.
6. Shall have alarms with audio and visual indicators for the following:
 - Skin high/ Low
 - Probe failure
 - Power failure, Heater Failure & system Failure too.
7. Accessories, spares and consumables
 - Temperature Probe: 5 Pcs
 - All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer.
8. Standards and Safety Requirements
 - Must submit ISO 13485:2003/AC:2007 for Medical Devices
 - Must submit Valid CE (93/42 EEC Directives) OR USFDA (510K) approved product certificate.
9. Warranty: 2 Years